

Instructions for Use

Version: 1.0.2

Revision date: 3-Jun-25



Free Radical Scavenging Capacity (DPPH) Assay Kit

Catalog No.: abx294424

Size: 96 tests

Detection Range: 17.39 $\mu\text{mol VC/L}$ – 125 $\mu\text{mol VC/L}$

Sensitivity: 17.39 $\mu\text{mol VC/L}$

Storage: Store all components at 4°C. Store the Chromogenic Reagent and Standards in the dark.

Application: For detection and quantification of Free radical scavenging capacity in serum, plasma, cell lysates and tissue homogenates.

Introduction

Abbexa's Free radical scavenging capacity Assay Kit is a quick, convenient, and sensitive method for measuring and calculating Free radical scavenging capacity. DPPH is a synthetic organic free radical which can be used to calculate antioxidant activity. The colored compound product of the reaction has an absorbance maximum at 525 nm. The intensity of the color is proportional to the free radical scavenging activity, which can then be calculated by comparing absorbance values before and after addition of sample antioxidants.

Kit components

1. 96-well microplate
2. Chromogenic Reagent: 1 Vial
3. Standard: 2 Vials
4. 15 ml Brown Plastic Bottle: 1
5. Plate sealer: 2

Materials required but not provided

1. Microplate reader (525 nm)
2. PBS (0.01 M, pH 7.4)
3. Ethanol (anhydrous)
4. Ethanol (80%)
5. Pipette and pipette tips
6. Sterile 2 ml microcentrifuge tubes
7. Ultrasonicator
8. Centrifuge

Instructions for Use

Version: 1.0.2

Revision date: 3-Jun-25

Protocol

A. Preparation of samples and reagents

1. Samples

Isolate the test samples soon after collecting and analyze immediately or aliquot and store at -20°C or -80°C for long-term storage. Avoid multiple freeze-thaw cycles.

The following sample preparation methods are intended as a guide and may be adjusted as required depending on the specific samples used.

- **Serum/Plasma:** Samples can be tested directly. Assay immediately or store at -80°C for up to one month.
- **Cell Lysate:** Harvest at least 1×10^6 cells and wash with PBS (0.01 M, pH 7.4). Add into 200 μ l of 80 % ethanol per 1×10^6 cells and homogenize manually by ultrasonication at 4°C. Centrifuge at $10,000 \times g$ for 10 minutes at 4°C. Collect the supernatant and store on ice for immediate detection.
- **Tissue Homogenates:** Carefully weigh at least 20 mg of tissue and wash with cold PBS (0.01 M, pH 7.4). Add into 180 μ l of 80 % ethanol per 20 mg of tissue sample and homogenize manually, using a mechanical homogenizer at 4°C. Centrifuge at $10,000 \times g$ for 10 minutes at 4°C. Collect the supernatant and store on ice for immediate detection.

It is recommended to carry out a preliminary experiment to determine the optimal dilution factor of samples before carrying out the formal experiment. Dilute samples into different concentrations with 80 % ethanol, then carry out the assay procedure. The recommended dilution factor for different samples is as follows (for reference only):

Sample Type	Dilution Factor
Horse serum	1
Dog plasma	1
Mouse serum	1
Rat serum	1
10 % Rat liver tissue homogenate	3-6
10 % Rat heart tissue homogenate	1
10 % Mouse heart tissue homogenate	1
10 % Cow liver tissue homogenate	4-8
10 % Mouse lung tissue homogenate	3-6
10 % Mouse stomach tissue homogenate	3-6
1.92×10^6 CHO cells lysate	1
2.164×10^6 Molt-4 cells lysate	1

Note:

- Fresh samples or recently obtained samples are recommended to prevent degradation and denaturalization that may lead to erroneous results.

Instructions for Use

Version: 1.0.2

Revision date: 3-Jun-25

- Lysis buffers may interfere with the kit. It is therefore recommended to use mechanical lysis methods for cell lysates and tissue homogenates.

2. Reagents

- Chromogenic Reagent Working Solution:** Dissolve a vial of Chromogenic Reagent with 1 ml of anhydrous ethanol and sonicate for 1 minute to mix thoroughly. Transfer the solution to the provided brown plastic bottle, then add 11 ml of anhydrous ethanol and sonicate for 5 minutes or incubate at room temperature for 2 hours in the dark. The prepared Chromogenic Working Solution can be stored at 4°C for up to 2 weeks in the dark.
- 12.5 mmol/L Standard Solution:** Dissolve a vial of Standard with 1 ml of anhydrous ethanol and mix thoroughly. The prepared solution can be stored at 4°C for up to 1 week in the dark.
- 125 µmol/L Standard Working Solution:** Mix 10 µl of 12.5 mmol/L Standard Solution with 990 µl of anhydrous ethanol and mix thoroughly. Prepare fresh before use and use immediately.
- Standard Dilutions:** Prepare standard dilutions as summarized in the following table:

Standard Dilution (µmol/L)	0	25	50	62.5	75	87.5	100	125
125 µmol/L Standard (µl)	0	40	80	100	120	140	160	200
Anhydrous ethanol (µl)	200	160	120	100	80	60	40	0

For the blank, or 0 µmol/L standard, use pure anhydrous ethanol. The volume of each standard will be 200 µl.

Note:

- Allow all reagents to equilibrate to room temperature before use.

B. Assay Procedure

Serum and Plasma samples

- Label microcentrifuge tubes for standards, samples, and controls. *Each sample requires a corresponding control. It is strongly recommended to prepare all the tubes in duplicate.*
- Add 400 µl of each standard dilution to the corresponding standard tubes.
- Add 400 µl of sample to the sample tubes.
- Add 400 µl of sample to the control tubes.
- Add 500 µl of Chromogenic Reagent Working Solution to the standard and sample tubes, and 500 µl of anhydrous ethanol to the control tubes.
- Mix thoroughly and incubate all tubes at room temperature for 10 minutes in the dark. Centrifuge at 4000 × g for 5 minutes at 4°C.
- Assign and record microplate well positions for each standard, sample, and control. Transfer 180 µl of supernatant from each tube to the corresponding microplate wells.
- Measure the OD of each well with a microplate reader at 525 nm.

Instructions for Use

Version: 1.0.2

Revision date: 3-Jun-25

Cell Lysate and Tissue Homogenate samples

1. Assign microplate wells for each standard, sample, and control. *Each sample requires a corresponding control. It is strongly recommended to prepare all the wells in duplicate.*
2. Add 80 µl of each standard dilution to the corresponding standard wells.
3. Add 80 µl of sample to the sample wells.
4. Add 80 µl of sample to the control wells.
5. Add 100 µl of Chromogenic Reagent Working Solution to the standard and sample wells, and 100 µl of anhydrous ethanol to the control wells.
6. Mix thoroughly and incubate at room temperature for 10 minutes in the dark.
7. Measure the OD of each well with a microplate reader at 525 nm.

C. Calculation of Results

Plot the standard curve, using the OD of the standard dilutions (adjusted for the blank) on the y-axis, and their respective concentrations on the x-axis. The curve should form a straight line, described by the formula $y = ax + b$. Based on this curve, the free radical scavenging capacity in each sample well can be derived with the following formulae:

$$\text{Free radical scavenging capacity } (\mu\text{mol VC/L}) = \frac{\text{OD}_{\text{Blank}} - (\text{OD}_{\text{Sample}} - \text{OD}_{\text{Control}}) - b}{a} \times f$$

$$\text{Free radical scavenging rate } (\%) = \frac{\text{OD}_{\text{Blank}} - (\text{OD}_{\text{Sample}} - \text{OD}_{\text{Control}})}{\text{OD}_{\text{Blank}}} \times 100 \%$$

where:

OD_{Blank}	OD value of 0 µmol/L blank standard
$\text{OD}_{\text{Sample}}$	OD value of sample
$\text{OD}_{\text{Control}}$	OD value of control
a	Gradient of the standard curve ($y = ax + b$)
b	Y-intercept of the standard curve ($y = ax + b$)
f	Dilution factor of sample before assay

Technical Support

For troubleshooting and technical assistance, please contact us at support@abbexa.com.