

Human Carbohydrate Sulfotransferase 15 (CHST15) ELISA Kit

Catalogue No.:abx386531

Human Carbohydrate Sulfotransferase 15 (CHST15) ELISA Kit is an ELISA Kit for the in vitro quantitative measurement of Human CHST15 concentrations in serum, plasma, tissue homogenates and other biological fluids.

Target:	Carbohydrate Sulfotransferase 15 (CHST15)
Reactivity:	Human
Tested Applications:	ELISA
Recommended dilutions	: Optimal dilutions/concentrations should be determined by the end user.
Storage:	Shipped at 4 °C. Upon receipt, store the kit according to the storage instruction in the kit's manual.
Validity:	The validity for this kit is at least 6 months. Up to 12 months validity can be provided on request.
Stability:	The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout.
UniProt Primary AC:	Q7LFX5 (<u>UniProt</u> , <u>ExPASy</u>)
Gene Symbol:	CHST15
GenelD:	51363
OMIM:	<u>608277</u>
HGNC:	18137
KEGG:	hsa:51363
Ensembl:	ENSG00000182022
String:	<u>9606.ENSP00000402394</u>
Test Range:	39.06 pg/ml - 2500 pg/ml
Sensitivity:	< 18.8 pg/ml

Datasheet Version: 3.0.0

Revision date: 02 Apr 2025



Standard Form:	Lyophilized
Detection Method:	Colorimetric
Assay Type:	Sandwich
Assay Data:	Quantitative
Sample Type:	Serum, plasma, tissue homogenates and other biological fluids.
Note:	THIS PRODUCT IS FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC OR THERAPEUTIC PROCEDURES. The range and sensitivity is subject to change. Please contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit. If you require a specific range, please contact us in advance or write your request in your order comments. Please note that our kits are optimised for detection of native samples, rather than recombinant proteins. We are unable to guarantee detection of recombinant proteins, as they may have different sequences or tertiary structures to the native protein.