Datasheet

Version: 3.0.0 Revision date: 04 Apr 2025



Human Magnesium transporter MRS2 homolog, mitochondrial (MRS2) ELISA Kit

Catalogue No.:abx533688

Human Magnesium transporter MRS2 homolog, mitochondrial (MRS2) ELISA Kit is an ELISA Kit for the in vitro quantitative measurement of Human Magnesium transporter MRS2 homolog, mitochondrial (MRS2) concentrations in tissue homogenates, cell lysates and other biological fluids.

Target: Magnesium transporter MRS2 homolog, mitochondrial (MRS2)

Reactivity: Human

Tested Applications: ELISA

Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Storage: Shipped at 4 °C. Upon receipt, store the kit according to the storage instruction in the kit's manual.

Validity: The validity for this kit is 6 months.

Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5% within

the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested

that the whole assay is performed by the same user throughout.

UniProt Primary AC: Q9HD23 (UniProt, ExPASy)

Gene Symbol: MRS2

KEGG: hsa:57380

Test Range: 78 pg/ml - 5000 pg/ml

Sensitivity: < 30 pg/ml

Standard Form: Lyophilized

Detection Method: Colorimetric

Assay Type: Sandwich

Assay Data: Quantitative

1 of 2

Website: www.abbexa.com · Email: info@abbexa.com

Datasheet

Version: 3.0.0 Revision date: 04 Apr 2025



Sample Type: Tissue homogenates, cell lysates and other biological fluids.

Note: THIS PRODUCT IS FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC OR

THERAPEUTIC PROCEDURES.

The range and sensitivity is subject to change. Please contact us for the latest product information. For accurate results, sample concentrations must be diluted to mid-range of the kit. If you require a specific range, please contact us in advance or write your request in your order comments. Please note that our kits are optimised for detection of native samples, rather than recombinant proteins. We are unable to guarantee detection of recombinant proteins, as they may have different sequences or tertiary structures to the native protein.