

## Western Enhanced Chemiluminescent (ECL) Substrate (Low Pico to Mid Femto Sensitivity)

Catalogue No.:abx299717

Western Enhanced Chemiluminescent (ECL) Substrate is a luminol-based chemiluminescent substrate for immunoblots with HRP-conjugated secondary antibodies and PDVF or nitrocellulose membranes. It is compatible with Western Blotting markers and is optimised for film and CCD-based imaging. This product has a high degree of sensitivity and enhanced chemiluminescence duration and is suitable for use with low picogram to mid femtogram detection of antigen. The 600 ml size contains 6 × 50 ml Peroxide and 6 × 50 ml Luminol.

Target:	Western Enhanced Chemiluminescent (ECL) Substrate (Low Pico to Mid Femto Sensitivity)
Tested Applications: WB	
Storage:	Store at room temperature for up to 12 months, or at 4 °C for up to 24 months. Avoid exposure to light.
Buffer:	Peroxide contains ≤ 1.2% sodium perborate tetrahydrate.
	Luminol contains $\leq$ 3.8% Tris and $\leq$ 0.13% hydrochloric acid.
Note:	THIS PRODUCT IS FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC, THERAPEUTIC OR COSMETIC PROCEDURES. NOT FOR HUMAN OR ANIMAL CONSUMPTION.
Directions for use:	<ol> <li>Keep the membrane moist in the wash buffer while preparing the substrate mixture. Please ensure the membrane does not dry out during the subsequent steps.</li> <li>Mix Luminol solution and Peroxide Solution in a 1:1 ratio, and thoroughly mix the chemiluminescent substrate solution well.</li> </ol>
	<ol> <li>Prepare 0.1 ml of solution per cm<sup>2</sup> of membrane. For a mini-sized membrane (7 × 8.5 cm), 5 ml of solution is sufficient. For a midi-sized membrane (8.5 × 13.5 cm), 10 ml of solution is sufficient.</li> <li>Place the membrane with the protein side up on a clear and level surface or in a clean container.</li> <li>Remove the membrane from the chemiluminescent substrate solution and drain off excessive solution.</li> <li>Place the membrane in a plastic sheet protector or in plastic wrap to prevent the membrane from drying.</li> <li>Image the membrane with a digital imager or by exposure to X-ray film.</li> </ol>